

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040818\
 Data File : VX000733.D
 Acq On : 08 Apr 2018 16:04
 Operator : JC/MD
 Sample : J2190-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16-040218

Quant Time: Apr 09 09:42:19 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040718W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 07 17:01:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	324278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	449383	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	419768	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	245305	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	164514	44.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.72%	
35) Dibromofluoromethane	5.51	113	137427	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
50) Toluene-d8	8.73	98	517279	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
62) 4-Bromofluorobenzene	11.15	95	212217	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
Target Compounds						
4) Vinyl Chloride	1.40	62	16431	6.23	ug/l	98
12) 1,1-Dichloroethene	2.38	96	3229	1.42	ug/l	99
16) Acetone	2.45	43	8245	7.09	ug/l	95
21) trans-1,2-Dichloroethene	3.17	96	4140	1.70	ug/l	94
27) cis-1,2-Dichloroethene	4.61	96	473535	152.44	ug/l	87
44) Trichloroethene	7.23	130	4208	1.28	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040818\
Data File : VX000733.D
Acq On : 08 Apr 2018 16:04
Operator : JC/MD
Sample : J2190-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16-040218

Quant Time: Apr 09 09:42:19 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040718W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 07 17:01:21 2018
Response via : Initial Calibration

