

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040818\
 Data File : VX000742.D
 Acq On : 08 Apr 2018 19:53
 Operator : JC/MD
 Sample : J2247-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CG-1

Quant Time: Apr 09 08:15:52 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040718W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 07 17:01:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	323301	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	450396	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	420228	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	248925	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	163767	44.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.60%	
35) Dibromofluoromethane	5.51	113	139677	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
50) Toluene-d8	8.72	98	518004	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
62) 4-Bromofluorobenzene	11.14	95	211320	46.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.22%	
Target Compounds						
16) Acetone	2.45	43	32929	28.39	ug/l	99
25) 2-Butanone	4.71	43	12319	6.07	ug/l	100
95) Naphthalene	13.84	128	10999	0.64	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040818\
Data File : VX000742.D
Acq On : 08 Apr 2018 19:53
Operator : JC/MD
Sample : J2247-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CG-1

Quant Time: Apr 09 08:15:52 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040718W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 07 17:01:21 2018
Response via : Initial Calibration

