


Data Path : W:\HPCHEM1\MSVOA X\DATA\VX040818\
Data File : VX000719.D
Acq On : 08 Apr 2018 09:15
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X040718W.M
Title : SW846 8260
Last Update : Sat Apr 07 17:01:21 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1723

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	8172	PASS
75	95	30	60	49.9	21767	PASS
95	95	100	100	100.0	43639	PASS
96	95	5	9	6.6	2866	PASS
173	174	0.00	2	1.0	399	PASS
174	95	50	100	90.4	39437	PASS
175	174	5	9	7.2	2849	PASS
176	174	95	101	97.2	38322	PASS
177	176	5	9	6.4	2471	PASS