

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041018\
 Data File : VX000762.D
 Acq On : 10 Apr 2018 21:36
 Operator : JC/MD
 Sample : J2157-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-040618-B

Quant Time: Apr 11 09:01:17 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	306553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	431714	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	404994	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	239348	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	159188	41.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.36%	
35) Dibromofluoromethane	5.51	113	133375	40.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.20%	
50) Toluene-d8	8.72	98	496501	40.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.18%	
62) 4-Bromofluorobenzene	11.14	95	202188	39.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.92%	
Target Compounds						
16) Acetone	2.45	43	5576	4.18	ug/l	99
20) Methylene Chloride	2.86	84	480166	154.88	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041018\
Data File : VX000762.D
Acq On : 10 Apr 2018 21:36
Operator : JC/MD
Sample : J2157-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-040618-B

Quant Time: Apr 11 09:01:17 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

