

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041118\
 Data File : VX000771.D
 Acq On : 11 Apr 2018 11:52
 Operator : JC/MD
 Sample : VX0411WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0411WBL01

Quant Time: Apr 11 13:59:53 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	310176	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	429053	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	398779	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	232962	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	156096	39.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.82%	
35) Dibromofluoromethane	5.51	113	131195	39.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.38%	
50) Toluene-d8	8.72	98	494169	40.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.30%	
62) 4-Bromofluorobenzene	11.14	95	199782	39.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.46%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041118\
Data File : VX000771.D
Acq On : 11 Apr 2018 11:52
Operator : JC/MD
Sample : VX0411WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0411WBL01

Quant Time: Apr 11 13:59:53 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

