

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041118\
 Data File : VX000776.D
 Acq On : 11 Apr 2018 14:14
 Operator : JC/MD
 Sample : J2280-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S2

Quant Time: Apr 11 15:01:55 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	315681	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	436295	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	408472	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	242727	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	159031	39.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.90%	
35) Dibromofluoromethane	5.51	113	134267	39.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.88%	
50) Toluene-d8	8.72	98	503720	40.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.50%	
62) 4-Bromofluorobenzene	11.14	95	202542	39.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.22%	
Target Compounds						
30) Chloroform	5.23	83	1601	0.27	ug/l	95
43) Isopropyl Acetate	6.48	43	2672	0.33	ug/l	96
95) Naphthalene	13.84	128	89918	5.09	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041118\
Data File : VX000776.D
Acq On : 11 Apr 2018 14:14
Operator : JC/MD
Sample : J2280-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S2

Quant Time: Apr 11 15:01:55 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

