

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041118\
 Data File : VX000780.D
 Acq On : 11 Apr 2018 15:56
 Operator : JC/MD
 Sample : PB108143TB 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108143TB

Quant Time: Apr 12 07:37:38 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	308721	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	424861	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	396332	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	233983	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	153786	39.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.02%	
35) Dibromofluoromethane	5.51	113	131473	40.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.34%	
50) Toluene-d8	8.72	98	491900	40.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.72%	
62) 4-Bromofluorobenzene	11.15	95	196479	38.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.94%	
Target Compounds						
16) Acetone	2.45	43	4692	3.50	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041118\
Data File : VX000780.D
Acq On : 11 Apr 2018 15:56
Operator : JC/MD
Sample : PB108143TB 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108143TB

Quant Time: Apr 12 07:37:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

