

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
 Data File : VX000802.D
 Acq On : 12 Apr 2018 15:15
 Operator : JC/MD
 Sample : J2190-10DL 4X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-116-040218DL

Quant Time: Apr 13 07:22:56 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	282673	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	360108	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	327309	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	208543	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	159947	44.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.76%	
35) Dibromofluoromethane	5.51	113	129975	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
50) Toluene-d8	8.73	98	499901	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
62) 4-Bromofluorobenzene	11.15	95	189310	44.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.58%	
Target Compounds						
4) Vinyl Chloride	1.41	62	3696	1.20	ug/l	96
27) cis-1,2-Dichloroethene	4.60	96	108320	33.36	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
Data File : VX000802.D
Acq On : 12 Apr 2018 15:15
Operator : JC/MD
Sample : J2190-10DL 4X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-116-040218DL

Quant Time: Apr 13 07:22:56 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

