

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
 Data File : VX000804.D
 Acq On : 12 Apr 2018 16:07
 Operator : JC/MD
 Sample : J2314-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CA-AQIDW01-20180409

Quant Time: Apr 13 07:25:28 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	281007	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	361820	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	323226	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	198204	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	157668	44.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.00%	
35) Dibromofluoromethane	5.51	113	129712	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
50) Toluene-d8	8.72	98	493939	47.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.18%	
62) 4-Bromofluorobenzene	11.15	95	186347	43.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.78%	
Target Compounds						
16) Acetone	2.45	43	3600	2.95	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
Data File : VX000804.D
Acq On : 12 Apr 2018 16:07
Operator : JC/MD
Sample : J2314-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CA-AQIDW01-20180409

Quant Time: Apr 13 07:25:28 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

