

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
 Data File : VX000807.D
 Acq On : 12 Apr 2018 17:23
 Operator : JC/MD
 Sample : J2312-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 F-4-09-18

Quant Time: Apr 13 07:30:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	276581	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	362412	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	325483	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	204201	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	158895	45.56	ug/l	0.00
Spiked Amount			Recovery	=	91.12%	
35) Dibromofluoromethane	5.51	113	130647	46.79	ug/l	0.00
Spiked Amount			Recovery	=	93.58%	
50) Toluene-d8	8.73	98	493647	47.48	ug/l	0.00
Spiked Amount			Recovery	=	94.96%	
62) 4-Bromofluorobenzene	11.15	95	187843	43.67	ug/l	0.00
Spiked Amount			Recovery	=	87.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	49824	41.43	ug/l	98
20) Methylene Chloride	2.86	84	4770	1.71	ug/l	93
25) 2-Butanone	4.72	43	33411	16.82	ug/l	100
27) cis-1,2-Dichloroethene	4.61	96	1053	0.33	ug/l	82
29) Tetrahydrofuran	5.18	42	7821	6.20	ug/l	98
30) Chloroform	5.22	83	2197	0.43	ug/l	95
52) Toluene	8.79	92	1711	0.23	ug/l	94
70) Styrene	10.72	104	2069	0.24	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
Data File : VX000807.D
Acq On : 12 Apr 2018 17:23
Operator : JC/MD
Sample : J2312-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
F-4-09-18

Quant Time: Apr 13 07:30:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

