

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
 Data File : VX000823.D
 Acq On : 13 Apr 2018 01:03
 Operator : JC/MD
 Sample : J2163-08 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-01-041118-D

Quant Time: Apr 13 08:02:38 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	274713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	360289	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	323911	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	207969	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	157090	45.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.70%	
35) Dibromofluoromethane	5.51	113	128364	46.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.48%	
50) Toluene-d8	8.72	98	494752	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
62) 4-Bromofluorobenzene	11.15	95	190191	44.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.96%	
Target Compounds						
16) Acetone	2.45	43	5241	4.39	ug/l	91
20) Methylene Chloride	2.86	84	34490	12.41	ug/l	98
95) Naphthalene	13.84	128	65301	4.32	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
Data File : VX000823.D
Acq On : 13 Apr 2018 01:03
Operator : JC/MD
Sample : J2163-08 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-01-041118-D

Quant Time: Apr 13 08:02:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

