

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
 Data File : VX000829.D
 Acq On : 13 Apr 2018 03:37
 Operator : JC/MD
 Sample : J2335-04 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S2

Quant Time: Apr 13 08:13:50 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	279366	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	367444	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	335925	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	219375	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	159821	45.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.74%	
35) Dibromofluoromethane	5.51	113	131151	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
50) Toluene-d8	8.72	98	503478	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
62) 4-Bromofluorobenzene	11.15	95	196689	45.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.20%	
Target Compounds						
16) Acetone	2.45	43	14364	11.82	ug/l	96
20) Methylene Chloride	2.87	84	4451	1.58	ug/l #	84
25) 2-Butanone	4.73	43	6388	3.18	ug/l	98
95) Naphthalene	13.84	128	192250	12.04	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
Data File : VX000829.D
Acq On : 13 Apr 2018 03:37
Operator : JC/MD
Sample : J2335-04 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S2

Quant Time: Apr 13 08:13:50 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

