

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
 Data File : VX000831.D
 Acq On : 13 Apr 2018 04:28
 Operator : JC/MD
 Sample : J2337-01 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-12002

Quant Time: Apr 13 08:16:09 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	285842	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	364570	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	327234	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199297	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	159539	44.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.54%	
35) Dibromofluoromethane	5.51	113	132372	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
50) Toluene-d8	8.73	98	507652	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
62) 4-Bromofluorobenzene	11.15	95	187362	43.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.60%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	6032	4.85	ug/l	96
20) Methylene Chloride	2.86	84	10335	3.58	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
Data File : VX000831.D
Acq On : 13 Apr 2018 04:28
Operator : JC/MD
Sample : J2337-01 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-12002

Quant Time: Apr 13 08:16:09 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

