

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
 Data File : VX000835.D
 Acq On : 13 Apr 2018 06:10
 Operator : JC/MD
 Sample : J2337-09 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR200036

Quant Time: Apr 13 08:21:14 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	282288	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	371341	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	348097	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	231398	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	162537	45.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.32%	
35) Dibromofluoromethane	5.51	113	134005	46.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.68%	
50) Toluene-d8	8.72	98	519135	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
62) 4-Bromofluorobenzene	11.15	95	205137	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
Target Compounds						
16) Acetone	2.45	43	6363	5.18	ug/l	100
17) Carbon Disulfide	2.57	76	11414	1.45	ug/l	100
20) Methylene Chloride	2.87	84	62607	21.93	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
Data File : VX000835.D
Acq On : 13 Apr 2018 06:10
Operator : JC/MD
Sample : J2337-09 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR200036

Quant Time: Apr 13 08:21:14 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

