

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
 Data File : VX000836.D
 Acq On : 13 Apr 2018 06:36
 Operator : JC/MD
 Sample : J2337-11 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR200029

Quant Time: Apr 13 08:25:20 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	275120	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	361286	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	326016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	200111	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	157220	45.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.64%	
35) Dibromofluoromethane	5.51	113	128461	46.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.30%	
50) Toluene-d8	8.72	98	499018	48.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.30%	
62) 4-Bromofluorobenzene	11.15	95	187348	43.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.38%	
Target Compounds						
16) Acetone	2.45	43	4990	4.17	ug/l	92
20) Methylene Chloride	2.86	84	88467	31.80	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
Data File : VX000836.D
Acq On : 13 Apr 2018 06:36
Operator : JC/MD
Sample : J2337-11 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR200029

Quant Time: Apr 13 08:25:20 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

