

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
 Data File : VX000837.D
 Acq On : 13 Apr 2018 07:01
 Operator : JC/MD
 Sample : J2337-13 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR200032

Quant Time: Apr 13 08:26:49 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	286123	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	377539	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	342136	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	221165	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	164564	45.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.22%	
35) Dibromofluoromethane	5.51	113	135329	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
50) Toluene-d8	8.72	98	516853	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
62) 4-Bromofluorobenzene	11.15	95	200241	44.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.38%	
Target Compounds						
16) Acetone	2.45	43	5242	4.21	ug/l	97
20) Methylene Chloride	2.86	84	3427	1.18	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041218\
Data File : VX000837.D
Acq On : 13 Apr 2018 07:01
Operator : JC/MD
Sample : J2337-13 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR200032

Quant Time: Apr 13 08:26:49 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

