

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041318\
 Data File : VX000854.D
 Acq On : 13 Apr 2018 17:16
 Operator : JC/MD
 Sample : J2335-02 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S1

Quant Time: Apr 14 05:29:48 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254733	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	336052	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	303468	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	197838	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.08	65	146389	45.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.16%	
35) Dibromofluoromethane	5.51	113	119623	46.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.40%	
50) Toluene-d8	8.72	98	460505	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
62) 4-Bromofluorobenzene	11.15	95	177957	44.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	8928	8.06	ug/l	97
20) Methylene Chloride	2.86	84	12858	4.99	ug/l	93
25) 2-Butanone	4.73	43	3501	1.91	ug/l	91
76) 1,2,3-Trichloropropane	11.15	75	84897	24.88	ug/l #	36
95) Naphthalene	13.84	128	102130	7.09	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041318\
Data File : VX000854.D
Acq On : 13 Apr 2018 17:16
Operator : JC/MD
Sample : J2335-02 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S1

Quant Time: Apr 14 05:29:48 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

