

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041318\
 Data File : VX000862.D
 Acq On : 13 Apr 2018 20:43
 Operator : JC/MD
 Sample : J2348-10 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 281

Quant Time: Apr 14 05:44:34 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	244655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	319982	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	290498	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	179552	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	141021	45.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.42%	
35) Dibromofluoromethane	5.51	113	114529	46.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.92%	
50) Toluene-d8	8.73	98	443843	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
62) 4-Bromofluorobenzene	11.15	95	167715	44.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.32%	
Target Compounds						
16) Acetone	2.45	43	25957	24.40	ug/l	98
18) Methyl Acetate	2.79	43	4872	1.77	ug/l	98
20) Methylene Chloride	2.86	84	19771	7.99	ug/l	98
95) Naphthalene	13.84	128	92766	7.10	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041318\
Data File : VX000862.D
Acq On : 13 Apr 2018 20:43
Operator : JC/MD
Sample : J2348-10 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
281

Quant Time: Apr 14 05:44:34 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

