


Data Path : W:\HPCHEM1\MSVOA X\Data\VX041318\
Data File : VX000840.D
Acq On : 13 Apr 2018 09:20
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
BFB

Integration File: RTEINT.P

Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Title : SW846 8260
Last Update : Tue Apr 10 15:49:28 2018



AutoFind: Scans 1730, 1731, 1732; Background Corrected with Scan 1720

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.6	8466	PASS
75	95	30	60	48.3	21984	PASS
95	95	100	100	100.0	45517	PASS
96	95	5	9	7.0	3187	PASS
173	174	0.00	2	1.1	464	PASS
174	95	50	100	92.1	41912	PASS
175	174	5	9	7.3	3042	PASS
176	174	95	101	98.4	41232	PASS
177	176	5	9	6.4	2640	PASS