

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
 Data File : VX000868.D
 Acq On : 14 Apr 2018 12:06
 Operator : JC/MD
 Sample : J2353-01 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS032MW008-180410

Quant Time: Apr 16 06:34:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	257795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	344414	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	309907	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	191756	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	149436	45.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.94%	
35) Dibromofluoromethane	5.51	113	120365	45.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.72%	
50) Toluene-d8	8.72	98	473174	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
62) 4-Bromofluorobenzene	11.15	95	175944	43.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.08%	
Target Compounds						
4) Vinyl Chloride	1.40	62	791540	282.04	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	267849	90.46	ug/l	89
65) Chlorobenzene	10.15	112	41593	5.34	ug/l	98
91) 1,2-Dichlorobenzene	12.40	146	13848	2.17	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
Data File : VX000868.D
Acq On : 14 Apr 2018 12:06
Operator : JC/MD
Sample : J2353-01 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS032MW008-180410

Quant Time: Apr 16 06:34:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

