

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
 Data File : VX000871.D
 Acq On : 14 Apr 2018 13:42
 Operator : JC/MD
 Sample : J2353-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW003-180410

Quant Time: Apr 16 06:41:28 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	253017	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	340651	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	311590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	206217	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	147495	46.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.46%	
35) Dibromofluoromethane	5.51	113	119603	45.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.14%	
50) Toluene-d8	8.72	98	465502	47.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.28%	
62) 4-Bromofluorobenzene	11.15	95	181262	44.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.66%	
Target Compounds						
3) Chloromethane	1.32	50	1815	0.74	ug/l	94
4) Vinyl Chloride	1.40	62	3281	1.19	ug/l	94
16) Acetone	2.45	43	4612	4.19	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	74617	25.68	ug/l	89
65) Chlorobenzene	10.15	112	33265	4.24	ug/l	99
91) 1,2-Dichlorobenzene	12.40	146	9686	1.41	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
Data File : VX000871.D
Acq On : 14 Apr 2018 13:42
Operator : JC/MD
Sample : J2353-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW003-180410

Quant Time: Apr 16 06:41:28 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

