

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
 Data File : VX000872.D
 Acq On : 14 Apr 2018 14:08
 Operator : JC/MD
 Sample : J2353-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW144-180410

Quant Time: Apr 16 06:43:37 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250284	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	330158	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	303699	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	200383	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	145527	46.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.22%	
35) Dibromofluoromethane	5.51	113	117151	46.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.12%	
50) Toluene-d8	8.72	98	460248	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.15	95	178425	45.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.06%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1388	0.51	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	22358	7.78	ug/l	88
40) Benzene	6.15	78	4706	0.47	ug/l	96
65) Chlorobenzene	10.15	112	46645	6.11	ug/l	99
91) 1,2-Dichlorobenzene	12.40	146	11985	1.80	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
Data File : VX000872.D
Acq On : 14 Apr 2018 14:08
Operator : JC/MD
Sample : J2353-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW144-180410

Quant Time: Apr 16 06:43:37 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

