

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
 Data File : VX000879.D
 Acq On : 14 Apr 2018 17:09
 Operator : JC/MD
 Sample : J2348-08 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 275

Quant Time: Apr 16 12:34:56 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	236556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	319508	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	289112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	181571	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	139724	46.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.68%	
35) Dibromofluoromethane	5.51	113	112679	45.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.54%	
50) Toluene-d8	8.73	98	438847	47.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.76%	
62) 4-Bromofluorobenzene	11.15	95	167063	44.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
Data File : VX000879.D
Acq On : 14 Apr 2018 17:09
Operator : JC/MD
Sample : J2348-08 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
275

Quant Time: Apr 16 12:34:56 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

