

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
 Data File : VX000880.D
 Acq On : 14 Apr 2018 17:35
 Operator : JC/MD
 Sample : J2348-06 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 155

Quant Time: Apr 16 12:35:42 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	242241	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	322763	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	299444	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	193618	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	141850	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	
35) Dibromofluoromethane	5.52	113	115043	46.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.52%	
50) Toluene-d8	8.73	98	444403	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
62) 4-Bromofluorobenzene	11.15	95	173674	45.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.68%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
Data File : VX000880.D
Acq On : 14 Apr 2018 17:35
Operator : JC/MD
Sample : J2348-06 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
155

Quant Time: Apr 16 12:35:42 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

