

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
 Data File : VX000885.D
 Acq On : 14 Apr 2018 19:44
 Operator : JC/MD
 Sample : J2353-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW029-180410

Quant Time: Apr 16 07:29:41 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	238140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	318953	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	298673	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	200560	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	139408	46.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.86%	
35) Dibromofluoromethane	5.51	113	113827	46.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.64%	
50) Toluene-d8	8.73	98	443509	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
62) 4-Bromofluorobenzene	11.15	95	175021	46.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.48%	
Target Compounds						
16) Acetone	2.45	43	13025	12.58	ug/l	99
65) Chlorobenzene	10.15	112	2669	0.36	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
Data File : VX000885.D
Acq On : 14 Apr 2018 19:44
Operator : JC/MD
Sample : J2353-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW029-180410

Quant Time: Apr 16 07:29:41 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

