

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
 Data File : VX000886.D
 Acq On : 14 Apr 2018 20:10
 Operator : JC/MD
 Sample : J2353-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW004-180410

Quant Time: Apr 16 07:32:57 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	238767	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	319112	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	293959	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	186001	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	141389	46.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.92%	
35) Dibromofluoromethane	5.51	113	114534	46.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.18%	
50) Toluene-d8	8.73	98	446455	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
62) 4-Bromofluorobenzene	11.15	95	167601	44.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.50%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	105157	38.34	ug/l	88
91) 1,2-Dichlorobenzene	12.40	146	2230	0.36	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041418\
Data File : VX000886.D
Acq On : 14 Apr 2018 20:10
Operator : JC/MD
Sample : J2353-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW004-180410

Quant Time: Apr 16 07:32:57 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

