

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041518\
 Data File : VX000901.D
 Acq On : 15 Apr 2018 17:02
 Operator : JC/MD
 Sample : PB108205ZHE#17 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108205ZHE#17 5X

Quant Time: Apr 16 08:23:40 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.68	168	275138	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	378564	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	350219	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	227184	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	165866	47.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.62%	
35) Dibromofluoromethane	5.51	113	133164	45.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.32%	
50) Toluene-d8	8.72	98	522962	48.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.32%	
62) 4-Bromofluorobenzene	11.14	95	203002	45.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.36%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041518\
Data File : VX000901.D
Acq On : 15 Apr 2018 17:02
Operator : JC/MD
Sample : PB108205ZHE#17 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108205ZHE#17 5X

Quant Time: Apr 16 08:23:40 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.68 min Scan# 835

Delta R.T. 0.01 min

Lab File: VX000901.D

Acq: 15 Apr 2018 17:02

Instrument :

MSVOA_X

ClientSampleId :


PB108205ZHE#17 5X

Tot Ion:168 Resp: 275138

Ion Ratio Lower Upper




168 100

99 46.0 38.3 57.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 47.81 ug/l

RT: 6.07 min Scan# 900

Delta R.T. -0.00 min

Lab File: VX000901.D


Acq: 15 Apr 2018 17:02

Tgt Ion: 65 Resp: 165866

Ion Ratio Lower Upper



65 100


67 51.6 0.0 103.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.87 min Scan# 1031
 Delta R.T. 0.01 min
 Lab File: VX000901.D
 Acq: 15 Apr 2018 17:02

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108205ZHE#17 5X

Tot Ion:114 Resp: 378564
 Ion Ratio Lower Upper
 114 100
 63 18.7 0.0 38.4
 88 14.8 0.0 31.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX
 Ion 87.90 (87.60 to 88.60): VX






#35
 Dibromofluoromethane
 Concen: 45.66 ug/l
 RT: 5.51 min Scan# 808
 Delta R.T. -0.00 min
 Lab File: VX000901.D
 Acq: 15 Apr 2018 17:02

Tgt Ion:113 Resp: 133164
 Ion Ratio Lower Upper
 113 100
 111 102.8 81.5 122.3
 192 25.7 20.4 30.6



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.16 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. -0.00 min

Lab File: VX000901.D

Acq: 15 Apr 2018 17:02

Instrument :

MSVOA_X

ClientSampleId :


PB108205ZHE#17 5X

Tot Ion: 98 Resp: 522962

Ion Ratio Lower Upper




98 100

100 65.0 52.6 79.0



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V



#62

4-Bromofluorobenzene

Concen: 45.18 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. -0.01 min

Lab File: VX000901.D

Acq: 15 Apr 2018 17:02


Tgt Ion: 95 Resp: 203002

Ion Ratio Lower Upper

95 100

174 98.1 0.0 198.4



176 95.6 0.0 194.4




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1564
Delta R.T. -0.00 min
Lab File: VX000901.D
Acq: 15 Apr 2018 17:02

Instrument :
MSVOA_X
ClientSampleId :
PB108205ZHE#17 5X

Tot Ion:117 Resp: 350219
Ion Ratio Lower Upper
117 100
82 51.5 41.9 62.9
119 32.8 26.0 39.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.09 min Scan# 1886
Delta R.T. -0.00 min
Lab File: VX000901.D
Acq: 15 Apr 2018 17:02

Tgt Ion:152 Resp: 227184
Ion Ratio Lower Upper
152 100
115 53.2 37.9 113.7
150 155.6 0.0 348.0

