

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041618\
 Data File : VX000915.D
 Acq On : 16 Apr 2018 13:42
 Operator : JC/MD
 Sample : VX0416MBL01
 Misc : 5.00g/5.0mL/100uL/5mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0416MBL01

Quant Time: Apr 16 16:28:14 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277706	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	371198	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	340967	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	220901	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	163731	46.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.52%	
35) Dibromofluoromethane	5.51	113	132767	46.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.84%	
50) Toluene-d8	8.72	98	515862	48.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.90%	
62) 4-Bromofluorobenzene	11.14	95	198567	45.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041618\
Data File : VX000915.D
Acq On : 16 Apr 2018 13:42
Operator : JC/MD
Sample : VX0416MBL01
Misc : 5.00g/5.0mL/100uL/5mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX0416MBL01

Quant Time: Apr 16 16:28:14 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

