

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041618\
 Data File : VX000924.D
 Acq On : 16 Apr 2018 17:32
 Operator : JC/MD
 Sample : J2260-02ME
 Misc : 7.48g/10mL/100uL/5mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-2-0-6-BELOW-GRADE(CW)ME

Quant Time: Apr 17 06:36:40 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	240984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	329709	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	310043	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	207421	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	142794	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
35) Dibromofluoromethane	5.51	113	110960	43.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.36%	
50) Toluene-d8	8.72	98	461157	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
62) 4-Bromofluorobenzene	11.15	95	186008	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
Target Compounds						
20) Methylene Chloride	2.85	84	1119	0.46	ug/l	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041618\
Data File : VX000924.D
Acq On : 16 Apr 2018 17:32
Operator : JC/MD
Sample : J2260-02ME
Misc : 7.48g/10mL/100uL/5mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-2-0-6-BELOW-GRADE(CW)ME

Quant Time: Apr 17 06:36:40 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

