

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041618\
 Data File : VX000937.D
 Acq On : 16 Apr 2018 23:04
 Operator : JC/MD
 Sample : PB108244ZHE#09 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108244ZHE#09

Quant Time: Apr 17 07:08:45 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	258427	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	360211	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	331684	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	214645	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	159998	49.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.20%	
35) Dibromofluoromethane	5.51	113	126850	45.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.42%	
50) Toluene-d8	8.72	98	502485	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
62) 4-Bromofluorobenzene	11.14	95	192566	45.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.08%	
Target Compounds						
16) Acetone	2.45	43	3751	3.34	ug/l	Qvalue 100
18) Methyl Acetate	2.78	43	5104	1.76	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041618\
Data File : VX000937.D
Acq On : 16 Apr 2018 23:04
Operator : JC/MD
Sample : PB108244ZHE#09 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108244ZHE#09

Quant Time: Apr 17 07:08:45 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

