

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041618\
 Data File : VX000941.D
 Acq On : 17 Apr 2018 01:36
 Operator : JC/MD
 Sample : VX0416WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0416WBL02

Quant Time: Apr 17 07:20:12 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	260640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	362039	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	332682	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	216860	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	159908	48.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.32%	
35) Dibromofluoromethane	5.51	113	125947	45.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.30%	
50) Toluene-d8	8.72	98	498793	48.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.06%	
62) 4-Bromofluorobenzene	11.14	95	191934	44.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.34%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX041618\
Data File : VX000941.D
Acq On : 17 Apr 2018 01:36
Operator : JC/MD
Sample : VX0416WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0416WBL02

Quant Time: Apr 17 07:20:12 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

