

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041618\
 Data File : VX000940.D
 Acq On : 17 Apr 2018 00:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 17 07:16:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.564	0.549	2.7	83	0.00
3 P	Chloromethane	0.484	0.498	-2.9	89	0.00
4 C	Vinyl Chloride	0.544	0.550	-1.1#	88	0.00
5 T	Bromomethane	0.250	0.317	-26.8#	123	0.00
6 T	Chloroethane	0.330	0.337	-2.1	93	0.00
7 T	Trichlorofluoromethane	0.821	0.849	-3.4	90	0.00
8 T	Diethyl Ether	0.307	0.329	-7.2	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.484	0.502	-3.7	93	0.00
10 T	Methyl Iodide	0.311	0.414	-33.1#	109	0.00
11 T	Tert butyl alcohol	0.112	0.097	13.4	75	0.00
12 CM	1,1-Dichloroethene	0.459	0.467	-1.7#	90	0.00
13 T	Acrolein	0.054	0.027	50.0#	44#	0.00
14 T	Allyl chloride	0.849	0.844	0.6	88	0.00
15 T	Acrylonitrile	0.239	0.258	-7.9	93	0.00
16 T	Acetone	0.217	0.207	4.6	84	0.00
17 T	Carbon Disulfide	1.390	1.143	17.8	74	0.00
18 T	Methyl Acetate	0.563	0.675	-19.9	104	0.00
19 T	Methyl tert-butyl Ether	1.529	1.651	-8.0	94	0.00
20 T	Methylene Chloride	0.506	0.520	-2.8	94	0.00
21 T	trans-1,2-Dichloroethene	0.506	0.506	0.0	90	0.00
22 T	Diisopropyl ether	1.559	1.761	-13.0	99	0.00
23 T	Vinyl Acetate	1.369	1.438	-5.0	93	0.00
24 P	1,1-Dichloroethane	0.875	0.952	-8.8	97	0.00
25 T	2-Butanone	0.359	0.361	-0.6	86	0.00
26 T	2,2-Dichloropropane	0.869	0.708	18.5	72	0.00
27 T	cis-1,2-Dichloroethene	0.574	0.608	-5.9	93	0.00
28 T	Bromochloromethane	0.352	0.352	0.0	92	0.00
29 T	Tetrahydrofuran	0.228	0.234	-2.6	89	0.00
30 C	Chloroform	0.926	1.001	-8.1#	94	0.00
31 T	Cyclohexane	0.831	0.813	2.2	88	0.00
32 T	1,1,1-Trichloroethane	0.883	0.904	-2.4	89	0.00
33 S	1,2-Dichloroethane-d4	0.630	0.634	-0.6	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.385	0.392	-1.8	90	0.00
36 T	1,1-Dichloropropene	0.551	0.547	0.7	91	0.00
37 T	Ethyl Acetate	0.538	0.535	0.6	88	0.00
38 T	Carbon Tetrachloride	0.616	0.587	4.7	85	0.00
39 T	Methylcyclohexane	0.661	0.615	7.0	86	0.00
40 TM	Benzene	1.530	1.588	-3.8	93	0.00
41 T	Methacrylonitrile	0.311	0.321	-3.2	92	0.00
42 TM	1,2-Dichloroethane	0.571	0.599	-4.9	94	0.00
43 T	Isopropyl Acetate	0.922	0.925	-0.3	90	0.00
44 TM	Trichloroethene	0.476	0.484	-1.7	91	0.00
45 C	1,2-Dichloropropane	0.390	0.415	-6.4#	94	0.00

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041618\
 Data File : VX000940.D
 Acq On : 17 Apr 2018 00:45
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 17 07:16:02 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.272	0.281	-3.3	93	0.00
47 T	Bromodichloromethane	0.548	0.550	-0.4	88	0.00
48 T	Methyl methacrylate	0.478	0.490	-2.5	91	0.00
49 T	1,4-Dioxane	0.010	0.008	20.0	70	0.00
50 S	Toluene-d8	1.434	1.444	-0.7	92	0.00
51 T	4-Methyl-2-Pentanone	0.541	0.566	-4.6	92	0.00
52 CM	Toluene	1.015	1.049	-3.3#	93	0.00
53 T	t-1,3-Dichloropropene	0.658	0.648	1.5	87	0.00
54 T	cis-1,3-Dichloropropene	0.640	0.618	3.4	84	0.00
55 T	1,1,2-Trichloroethane	0.403	0.429	-6.5	97	0.00
56 T	Ethyl methacrylate	0.622	0.639	-2.7	91	0.00
57 T	1,3-Dichloropropane	0.641	0.693	-8.1	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.290	0.319	-10.0	95	0.00
59 T	2-Hexanone	0.418	0.418	0.0	87	0.00
60 T	Dibromochloromethane	0.482	0.478	0.8	86	0.00
61 T	1,2-Dibromoethane	0.428	0.454	-6.1	94	0.00
62 S	4-Bromofluorobenzene	0.593	0.587	1.0	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.489	0.519	-6.1	98	0.00
65 PM	Chlorobenzene	1.258	1.294	-2.9	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.473	0.477	-0.8	92	0.00
67 C	Ethyl Benzene	2.097	2.137	-1.9#	93	0.00
68 T	m/p-Xylenes	0.828	0.848	-2.4	94	0.00
69 T	o-Xylene	0.798	0.823	-3.1	94	0.00
70 T	Styrene	1.345	1.391	-3.4	94	0.00
71 P	Bromoform	0.438	0.402	8.2	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.172	3.279	-3.4	94	0.00
74 T	N-amyl acetate	1.428	1.408	1.4	90	0.00
75 P	1,1,2,2-Tetrachloroethane	0.921	0.950	-3.1	94	0.00
76 T	1,2,3-Trichloropropane	0.862	0.868	-0.7	90	0.00
77 T	Bromobenzene	0.930	0.950	-2.2	94	0.00
78 T	n-propylbenzene	3.628	3.654	-0.7	93	0.00
79 T	2-Chlorotoluene	2.144	2.195	-2.4	94	0.00
80 T	1,3,5-Trimethylbenzene	2.728	2.756	-1.0	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.313	0.251	19.8	72	0.00
82 T	4-Chlorotoluene	2.588	2.621	-1.3	94	0.00
83 T	tert-Butylbenzene	2.791	2.820	-1.0	91	0.00
84 T	1,2,4-Trimethylbenzene	2.816	2.855	-1.4	93	0.00
85 T	sec-Butylbenzene	3.292	3.269	0.7	90	0.00
86 T	p-Isopropyltoluene	3.070	3.029	1.3	90	0.00
87 T	1,3-Dichlorobenzene	1.721	1.763	-2.4	94	0.00
88 T	1,4-Dichlorobenzene	1.763	1.792	-1.6	94	0.00
89 T	n-Butylbenzene	2.710	2.618	3.4	88	0.00

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041618\
Data File : VX000940.D
Acq On : 17 Apr 2018 00:45
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 17 07:16:02 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.520	0.457	12.1	78	0.00
91 T	1,2-Dichlorobenzene	1.661	1.737	-4.6	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.231	0.198	14.3	79	0.00
93 T	1,2,4-Trichlorobenzene	1.417	1.409	0.6	91	0.00
94 T	Hexachlorobutadiene	0.697	0.581	16.6	77	0.00
95 T	Naphthalene	3.638	3.675	-1.0	91	0.00
96 T	1,2,3-Trichlorobenzene	1.377	1.375	0.1	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6