

Data File : VX000970.D
Acq On : 17 Apr 2018 17:42
Operator : JC/MD
Sample : J2364-05
Misc : 6.54g/5.0mL/100uL/5mL//MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-B

Quant Time: Apr 18 07:31:15 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	231867	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	323002	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	305670	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	205130	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	143661	49.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.28%	
35) Dibromofluoromethane	5.51	113	108739	43.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.40%	
50) Toluene-d8	8.72	98	451100	48.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.36%	
62) 4-Bromofluorobenzene	11.15	95	181783	47.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.84%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000970.D
Acq On : 17 Apr 2018 17:42
Operator : JC/MD
Sample : J2364-05
Misc : 6.54g/5.0mL/100uL/5mL//MSVOA_X/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-B

Quant Time: Apr 18 07:31:15 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

