

Data File : VX000971.D
Acq On : 17 Apr 2018 18:08
Operator : JC/MD
Sample : J2314-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CA-AQIDW01-20180409

Quant Time: Apr 17 18:28:57 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	255568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	350889	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	329710	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	217437	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	160748	49.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.76%	
35) Dibromofluoromethane	5.51	113	123070	45.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.04%	
50) Toluene-d8	8.72	98	496421	49.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.64%	
62) 4-Bromofluorobenzene	11.14	95	192718	46.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.56%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX000971.D
Acq On : 17 Apr 2018 18:08
Operator : JC/MD
Sample : J2314-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CA-AQIDW01-20180409

Quant Time: Apr 17 18:28:57 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

