

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041718\
 Data File : VX000974.D
 Acq On : 17 Apr 2018 19:24
 Operator : JC/MD
 Sample : PB108300TB 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108300TB 5X

Quant Time: Apr 18 05:25:47 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251329	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	343594	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	315843	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	202489	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	154892	48.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.76%	
35) Dibromofluoromethane	5.51	113	121118	45.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.50%	
50) Toluene-d8	8.72	98	481911	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
62) 4-Bromofluorobenzene	11.14	95	182684	44.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.60%	
Target Compounds						
16) Acetone	2.45	43	3528	3.23	ug/l	99
18) Methyl Acetate	2.79	43	4348	1.54	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041718\
Data File : VX000974.D
Acq On : 17 Apr 2018 19:24
Operator : JC/MD
Sample : PB108300TB 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108300TB 5X

Quant Time: Apr 18 05:25:47 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

