

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041718\
 Data File : VX000976.D
 Acq On : 17 Apr 2018 20:15
 Operator : JC/MD
 Sample : J2410-08 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T-3

Quant Time: Apr 18 05:48:38 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254232	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	352638	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	322551	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	202206	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	159549	49.77	ug/l	0.00
Spiked Amount						
			Recovery	=	99.54%	
35) Dibromofluoromethane	5.51	113	123130	45.32	ug/l	0.00
Spiked Amount						
			Recovery	=	90.64%	
50) Toluene-d8	8.72	98	488150	48.26	ug/l	0.00
Spiked Amount						
			Recovery	=	96.52%	
62) 4-Bromofluorobenzene	11.14	95	184978	44.20	ug/l	0.00
Spiked Amount						
			Recovery	=	88.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	4232	3.83	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041718\
Data File : VX000976.D
Acq On : 17 Apr 2018 20:15
Operator : JC/MD
Sample : J2410-08 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T-3

Quant Time: Apr 18 05:48:38 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

