

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041718\
 Data File : VX000977.D
 Acq On : 17 Apr 2018 20:40
 Operator : JC/MD
 Sample : J2411-03 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DRUM-1-3

Quant Time: Apr 18 05:50:46 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251320	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	349828	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	322826	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	204384	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	158892	50.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.28%	
35) Dibromofluoromethane	5.51	113	123652	45.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.76%	
50) Toluene-d8	8.72	98	488298	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
62) 4-Bromofluorobenzene	11.14	95	185701	44.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.46%	
Target Compounds						
16) Acetone	2.45	43	14579	13.34	ug/l	98
18) Methyl Acetate	2.78	43	4989	1.76	ug/l	99
20) Methylene Chloride	2.86	84	53751	21.15	ug/l	97
25) 2-Butanone	4.71	43	6045	3.35	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041718\
Data File : VX000977.D
Acq On : 17 Apr 2018 20:40
Operator : JC/MD
Sample : J2411-03 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRUM-1-3

Quant Time: Apr 18 05:50:46 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

