

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041718\
 Data File : VX000980.D
 Acq On : 17 Apr 2018 21:56
 Operator : JC/MD
 Sample : PB108300ZHE#18 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB108300ZHE#18 5X

Quant Time: Apr 18 08:11:01 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 10 15:49:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	251782	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	350670	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	321329	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	209066	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	156999	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
35) Dibromofluoromethane	5.51	113	122021	45.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.32%	
50) Toluene-d8	8.72	98	486997	48.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.82%	
62) 4-Bromofluorobenzene	11.14	95	187000	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\Data\VX041718\
Data File : VX000980.D
Acq On : 17 Apr 2018 21:56
Operator : JC/MD
Sample : PB108300ZHE#18 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB108300ZHE#18 5X

Quant Time: Apr 18 08:11:01 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041018W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 10 15:49:28 2018
Response via : Initial Calibration

