

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042018\
 Data File : VX001018.D
 Acq On : 20 Apr 2018 19:16
 Operator : JC/MD
 Sample : J2459-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-180418

Quant Time: Apr 21 06:20:37 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	229281	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	318701	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	292661	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	182826	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	149501	51.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.88%	
35) Dibromofluoromethane	5.51	113	113245	46.60	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.20%	
50) Toluene-d8	8.72	98	446685	49.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.42%	
62) 4-Bromofluorobenzene	11.14	95	167731	44.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.96%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042018\
Data File : VX001018.D
Acq On : 20 Apr 2018 19:16
Operator : JC/MD
Sample : J2459-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-180418

Quant Time: Apr 21 06:20:37 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

