

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042018\
 Data File : VX001019.D
 Acq On : 20 Apr 2018 19:41
 Operator : JC/MD
 Sample : J2459-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-180418

Quant Time: Apr 21 06:18:56 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232016	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	328011	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	306387	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	201549	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	150550	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
35) Dibromofluoromethane	5.51	113	116072	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
50) Toluene-d8	8.72	98	460500	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
62) 4-Bromofluorobenzene	11.14	95	179364	46.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.42%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	9758	3.48	ug/l	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042018\
Data File : VX001019.D
Acq On : 20 Apr 2018 19:41
Operator : JC/MD
Sample : J2459-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-180418

Quant Time: Apr 21 06:18:56 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

