

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042018\
 Data File : VX001035.D
 Acq On : 21 Apr 2018 03:18
 Operator : JC/MD
 Sample : J2377-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0442-180412

Quant Time: Apr 23 12:52:32 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	246080	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	348431	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	323510	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	214055	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161203	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
35) Dibromofluoromethane	5.51	113	123294	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
50) Toluene-d8	8.72	98	485863	49.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.92%	
62) 4-Bromofluorobenzene	11.14	95	189294	45.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.82%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	991	0.33	ug/l	74
40) Benzene	6.15	78	2554	0.22	ug/l	95
65) Chlorobenzene	10.15	112	137428	15.95	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042018\
Data File : VX001035.D
Acq On : 21 Apr 2018 03:18
Operator : JC/MD
Sample : J2377-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0442-180412

Quant Time: Apr 23 12:52:32 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

