

Data File : VX001038.D
Acq On : 21 Apr 2018 04:34
Operator : JC/MD
Sample : J2377-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW026-180412

Quant Time: Apr 21 06:45:57 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232906	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	329298	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	304717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	188718	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	153603	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
35) Dibromofluoromethane	5.51	113	116993	46.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.18%	
50) Toluene-d8	8.72	98	464585	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
62) 4-Bromofluorobenzene	11.14	95	175153	44.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
27) cis-1,2-Dichloroethene	4.60	96	1202	0.43	ug/l	82
44) Trichloroethene	7.22	130	8481	2.58	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX001038.D
Acq On : 21 Apr 2018 04:34
Operator : JC/MD
Sample : J2377-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW026-180412

Quant Time: Apr 21 06:45:57 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

