

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042018\
 Data File : VX001042.D
 Acq On : 21 Apr 2018 06:16
 Operator : JC/MD
 Sample : J2377-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014MW028-180412

Manual Integrations
 APPROVED

MMDadoda
 4/23/2018 1:27:47 PM

Quant Time: Apr 23 12:51:31 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	230663	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	324066	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	307614	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	206897	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	152781	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
35) Dibromofluoromethane	5.51	113	118644	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
50) Toluene-d8	8.72	98	460982	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.14	95	179682	46.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.72%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	1003	0.36	ug/l	77
40) Benzene	6.15	78	2234	0.21	ug/l #	79
65) Chlorobenzene	10.15	112	139334	17.01	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042018\
Data File : VX001042.D
Acq On : 21 Apr 2018 06:16
Operator : JC/MD
Sample : J2377-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW028-180412

Manual Integrations
APPROVED

MMDadoda
4/23/2018 1:27:47 PM

Quant Time: Apr 23 12:51:31 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

