

Data File : VX001043.D
Acq On : 21 Apr 2018 06:41
Operator : JC/MD
Sample : J2377-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014RW032-180412

Quant Time: Apr 23 00:55:48 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	237527	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	331532	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	314322	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	204989	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	153864	51.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.20%	
35) Dibromofluoromethane	5.51	113	118009	46.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.36%	
50) Toluene-d8	8.72	98	468489	50.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.24%	
62) 4-Bromofluorobenzene	11.14	95	183939	46.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX001043.D

Acq On : 21 Apr 2018 06:41

Operator : JC/MD

Sample : J2377-03

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 47 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

LF014RW032-180412

Quant Time: Apr 23 00:55:48 2018

Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M

Quant Title : SW846 8260

QLast Update : Fri Apr 20 05:03:28 2018

Response via : Initial Calibration

