

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
 Data File : VX001057.D
 Acq On : 23 Apr 2018 14:37
 Operator : JC/MD
 Sample : J2377-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014RW131-180412

Quant Time: Apr 24 07:30:09 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	223529	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	314206	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	287858	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	186084	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	146172	51.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.18%	
35) Dibromofluoromethane	5.51	113	112767	47.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.14%	
50) Toluene-d8	8.72	98	441393	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
62) 4-Bromofluorobenzene	11.14	95	166224	44.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.42%	
Target Compounds						
4) Vinyl Chloride	1.40	62	25564	9.55	ug/l	99
12) 1,1-Dichloroethene	2.38	96	1263	0.58	ug/l	98
27) cis-1,2-Dichloroethene	4.61	96	285004	105.53	ug/l	89
44) Trichloroethene	7.22	130	5025	1.60	ug/l	92
65) Chlorobenzene	10.15	112	92988	12.13	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
Data File : VX001057.D
Acq On : 23 Apr 2018 14:37
Operator : JC/MD
Sample : J2377-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014RW131-180412

Quant Time: Apr 24 07:30:09 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

