

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
 Data File : VX001058.D
 Acq On : 23 Apr 2018 15:02
 Operator : JC/MD
 Sample : J2377-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014MW006-180412

Quant Time: Apr 24 07:31:37 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	222002	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	312465	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	296597	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	201899	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	145776	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
35) Dibromofluoromethane	5.51	113	110547	46.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.80%	
50) Toluene-d8	8.72	98	440260	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	11.14	95	174605	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
Target Compounds						
4) Vinyl Chloride	1.40	62	2467	0.93	ug/l	98
27) cis-1,2-Dichloroethene	4.61	96	34861	13.00	ug/l	89
65) Chlorobenzene	10.15	112	8342	1.06	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
Data File : VX001058.D
Acq On : 23 Apr 2018 15:02
Operator : JC/MD
Sample : J2377-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW006-180412

Quant Time: Apr 24 07:31:37 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

