

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
 Data File : VX001060.D
 Acq On : 23 Apr 2018 15:53
 Operator : JC/MD
 Sample : J2377-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014MW015-180412

Quant Time: Apr 24 07:35:45 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	216485	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	307259	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	285058	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	189217	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	142553	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
35) Dibromofluoromethane	5.51	113	109846	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
50) Toluene-d8	8.72	98	429624	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
62) 4-Bromofluorobenzene	11.15	95	166553	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
Target Compounds						
4) Vinyl Chloride	1.40	62	144405	55.73	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	107532	41.11	ug/l	88
40) Benzene	6.15	78	7579	0.76	ug/l	94
52) Toluene	8.79	92	3326	0.51	ug/l	99
65) Chlorobenzene	10.15	112	160552	21.15	ug/l	98
88) 1,4-Dichlorobenzene	12.10	146	14059	2.04	ug/l	87
91) 1,2-Dichlorobenzene	12.40	146	9771	1.50	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
Data File : VX001060.D
Acq On : 23 Apr 2018 15:53
Operator : JC/MD
Sample : J2377-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW015-180412

Quant Time: Apr 24 07:35:45 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

