

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
 Data File : VX001061.D
 Acq On : 23 Apr 2018 16:19
 Operator : JC/MD
 Sample : J2464-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1TB09-180417

Quant Time: Apr 24 07:52:28 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	211102	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	293808	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	264006	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152431	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	137095	51.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.46%	
35) Dibromofluoromethane	5.51	113	105877	47.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.52%	
50) Toluene-d8	8.72	98	409517	49.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.86%	
62) 4-Bromofluorobenzene	11.14	95	146039	42.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.02%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
Data File : VX001061.D
Acq On : 23 Apr 2018 16:19
Operator : JC/MD
Sample : J2464-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1TB09-180417

Quant Time: Apr 24 07:52:28 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

