

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
 Data File : VX001062.D
 Acq On : 23 Apr 2018 16:44
 Operator : JC/MD
 Sample : J2464-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1TB08-180416

Quant Time: Apr 24 07:54:16 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	205153	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	287590	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	257580	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	149008	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.08	65	135065	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
35) Dibromofluoromethane	5.51	113	104037	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
50) Toluene-d8	8.72	98	400527	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
62) 4-Bromofluorobenzene	11.14	95	142482	41.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.74%	
Target Compounds						
16) Acetone	2.45	43	3140	3.21	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
Data File : VX001062.D
Acq On : 23 Apr 2018 16:44
Operator : JC/MD
Sample : J2464-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1TB08-180416

Quant Time: Apr 24 07:54:16 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

