

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
 Data File : VX001063.D
 Acq On : 23 Apr 2018 17:10
 Operator : JC/MD
 Sample : J2464-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF017MW001-180417

Quant Time: Apr 24 08:00:49 2018
 Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 20 05:03:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	205524	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	287839	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	250962	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	135953	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	136210	52.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.56%	
35) Dibromofluoromethane	5.51	113	104637	47.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.34%	
50) Toluene-d8	8.72	98	397942	49.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.06%	
62) 4-Bromofluorobenzene	11.14	95	134601	39.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.04%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA X\DATA\VX042318\
Data File : VX001063.D
Acq On : 23 Apr 2018 17:10
Operator : JC/MD
Sample : J2464-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF017MW001-180417

Quant Time: Apr 24 08:00:49 2018
Quant Method : W:\HPCHEM1\MSVOA_X\METHOD\82X041918W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 20 05:03:28 2018
Response via : Initial Calibration

